

**Test Report No. 7191153239-EEC16/01-CSL**  
**dated 20 Dec 2016**



PSB Singapore

**Choose certainty.**  
**Add value.**

**Note:** This report is issued subject to the Testing and Certification Regulations of the TÜV SÜD Group and the General Terms and Conditions of Business of TÜV SÜD PSB Pte Ltd. In addition, this report is governed by the terms set out within this report.

**SUBJECT:**

Testing of Vinyl Gloves submitted by Shandong Intco Medical Products Co., Ltd.  
on 19 Dec 2016.

**TESTED FOR:**

Shandong Intco Medical Products Co., Ltd.  
No. 9888, Qiwang Road,  
Naoshan Industry Park,  
Qingzhou, Shandong,  
China.

**TEST DATE:**

20 Dec 2016

**DESCRIPTION OF SAMPLES:**

| S/N | Product Description | Colour | Size | Lot no.             | Sample received | Manufacturer                              |
|-----|---------------------|--------|------|---------------------|-----------------|---|
| 1   | Vinyl Gloves        | Clear  | M    | -<br>(see remark 1) | 249 pieces      | Shandong Intco Medical Products Co., Ltd. |

Lot size as specified by client: 35,001 to 150,000 pieces

**METHOD OF TEST:**

ASTM D5250-06(2015) Standard Specification for Poly(vinyl chloride) Gloves for Medical Application  
- Clauses 6.1.2 & 7.3 Freedom from Holes, with reference to ASTM D5151-06(2015)  
- Inspection level G-I, AQL 2.5



Laboratory:  
TÜV SÜD PSB Pte. Ltd.  
No.1 Science Park Drive  
Singapore 118221

Phone : +65-6885 1333  
Fax : +65-6776 8670  
E-mail: enquiries@tuv-sud-psb.sg  
www.tuv-sud-psb.sg  
Co. Reg : 199002667R

Regional Head Office:  
TÜV SÜD Asia Pacific Pte. Ltd.  
1 Science Park Drive, #02-01  
Singapore 118221  
**TUV®**

**RESULTS:**


Sample: Vinyl Gloves, Clear, Size M

Table: Result for ASTM D5250-06(2015) Clauses 6.1.2 & 7.3

| Test               | Requirements   | No. of non-compliers allowed (pieces) | Number tested (pieces) | Actual no. of non-compliers found (pieces) | Inferred results |
|--------------------|----------------|---------------------------------------|------------------------|--|------------------|
| Freedom from holes | Shall not leak | 10                                    | 200                    | 4  | Passed           |

**REMARKS:**

1. The manufacturing batch code was not provided by the client.



Shareen Chan  
Engineer



Wong Bee Hui  
Product Manager  
Medical Health Services (NAM)

**APPENDIX:**



Photo : Vinyl Gloves, Clear, Size M

Please note that this Report is issued under the following terms :

1. This report applies to the sample of the specific product/equipment given at the time of its testing/calibration. The results are not used to indicate or imply that they are applicable to other similar items. In addition, such results must not be used to indicate or imply that TÜV SÜD PSB approves, recommends or endorses the manufacturer, supplier or user of such product/equipment, or that TÜV SÜD PSB in any way "guarantees" the later performance of the product/equipment. Unless otherwise stated in this report, no tests were conducted to determine long term effects of using the specific product/equipment.
2. The sample/s mentioned in this report is/are submitted/supplied/manufactured by the Client. TÜV SÜD PSB therefore assumes no responsibility for the accuracy of information on the brand name, model number, origin of manufacture, consignment or any information supplied.
3. Nothing in this report shall be interpreted to mean that TÜV SÜD PSB has verified or ascertained any endorsement or marks from any other testing authority or bodies that may be found on that sample.
4. This report shall not be reproduced wholly or in parts and no reference shall be made by the Client to TÜV SÜD PSB or to the report or results furnished by TÜV SÜD PSB in any advertisements or sales promotion.
5. Unless otherwise stated, the tests were carried out in TÜV SÜD PSB Pte Ltd, No.1 Science Park Drive Singapore 118221.

July 2011

